



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

**Customer:** NW Natural Goods  
**Product identity:** BEV-BO 021021-1  
**Client/Metric ID:** .  
**Laboratory ID:** 21-000940-0001

**Sample Date:** 01/26/21

### Summary

**Potency:**

Analyte per 355ml	Result	Limits	Units	Status	
CBD per 355ml	26.0		mg/355ml		CBD-Total per 355ml 26.0 mg/355ml
					THC-Total per 355ml <LOQ
(Reported in milligrams per serving)					

**Residual Solvents:**

*All analytes passing and less than LOQ.*

**Pesticides:**

Analyte	Result (mg/kg)	Limits (mg/kg)	Status
Multi-Residue Pesticide Profile <sup>1</sup>	< LOQ for all analytes		

**Metals:**

*Less than LOQ for all analytes.*

**Microbiology:**

*Less than LOQ for all analytes.*

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Testing in accordance with: OAR 333-007-0430



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794

**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10



**Customer:** NW Natural Goods

**Product identity:** BEV-BO 021021-1  
**Client/Metric ID:** .  
**Sample Date:** 01/26/21  
**Laboratory ID:** 21-000940-0001  
**Relinquished by:** Brian Ramos  
**Temp:** 13.3 °C  
**Serving Size #1:** 362.1 g  
**Density:** 1.020 g/ml

### Sample Results

Potency per 355ml		Method J AOAC 2015 V98-6 (mod)		Batch: 2100943		Analyze: 2/1/21 12:23:00 PM	
Analyte	Result	Limits	Units	LOQ	Notes		
CBC per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBC-A per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBC-Total per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.669			
CBD per 355ml	26.0		mg/355ml	0.356			
CBD-A per 355ml	< LOQ		mg/355ml	0.356			
CBD-Total per 355ml	26.0		mg/355ml	0.669			
CBDV per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBDV-A per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBDV-Total per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.665			
CBG per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBG-A per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBG-Total per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.665			
CBL per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
CBN per 355ml	< LOQ		mg/355ml	0.356			
Δ8-THC per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
Δ9-THC per 355ml	< LOQ		mg/355ml	0.356			
THC-A per 355ml	< LOQ		mg/355ml	0.356			
THC-Total per 355ml	< LOQ		mg/355ml	0.669			
THCV per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
THCV-A per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.356			
THCV-Total per 355ml <sup>†</sup>	< LOQ		mg/355ml	0.669			
<b>Total Cannabinoids 355ml<sup>†</sup></b>	<b>26.0</b>		mg/355ml				

### Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Notes
E.coli	< LOQ		cfu/g	10	2100821	01/30/21	AOAC 991.14 (Petrifilm)	X
Total Coliforms	< LOQ		cfu/g	10	2100821	01/30/21	AOAC 991.14 (Petrifilm)	X
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2100824	01/30/21	AOAC 2014.05 (RAPID)	X
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2100824	01/30/21	AOAC 2014.05 (RAPID)	X



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794

**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10



Solvents		Method EPA5021A				Units µg/g	Batch 2100841	Analyze 01/28/21 01:43 PM			
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethyl acetate	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether	< LOQ	5000	200	pass	
Ethylene glycol	< LOQ	620	200	pass		Ethylene oxide	< LOQ	50.0	30.0	pass	
Hexanes (sum)	< LOQ	290	150	pass		Isopropyl acetate	< LOQ	5000	200	pass	
Isopropylbenzene	< LOQ	70.0	30.0	pass		m,p-Xylene	< LOQ		200		
Methanol	< LOQ	3000	200	pass		Methylene chloride	< LOQ	600	200	pass	
Methylpropane	< LOQ		200			n-Butane	< LOQ		200		
n-Heptane	< LOQ	5000	200	pass		n-Hexane	< LOQ		30.0		
n-Pentane	< LOQ		200			o-Xylene	< LOQ		200		
Pentanes (sum)	< LOQ	5000	600	pass		Propane	< LOQ	5000	200	pass	
Tetrahydrofuran	< LOQ	720	100	pass		Toluene	< LOQ	890	100	pass	
Total Xylenes	< LOQ		400			Total Xylenes and Ethyl	< LOQ	2170	600	pass	

Pesticides		Method AOAC 2007.01 & EN 15662 (mod)			Units mg/kg	Batch 2100950	Analyze 02/01/21 04:25 PM			
Analyte	Result	Limits	Status	Notes						
Multi-Residue Pesticide Profile <sup>1</sup>	< LOQ for all analytes									

Metals									
Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Notes	
Arsenic	< LOQ		mg/kg	0.00197	2100857	01/28/21	AOAC 2013.06 (mod.)	X	
Cadmium	< LOQ		mg/kg	0.00197	2100857	01/28/21	AOAC 2013.06 (mod.)	X	
Lead	< LOQ		mg/kg	0.00197	2100857	01/28/21	AOAC 2013.06 (mod.)	X	
Mercury	< LOQ		mg/kg	0.00098	2100857	01/28/21	AOAC 2013.06 (mod.)	X	

Nutrition									
Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Notes	
Moisture (Loss on Drying)	99.4		g/100g	0.10	2100896	01/29/21	AOAC 925.10 (mod.)	X	
Water Activity	0.991		Aw	0.030	2100912	01/30/21	AOAC 978.18	X	



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

**Abbreviations**

**Limits:** Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220

**Limit(s) of Quantitation (LOQ):** The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

† = Analyte not NELAP accredited.

**Units of Measure**

cfu/g = Colony forming units per gram

g = Gram

g/ml = Gram per milliliter

g/100g = Grams per 100 Grams

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/355ml = Milligram per 355ml

% = Percentage of sample

Aw = Water Activity

% wt = µg/g divided by 10,000

**Glossary of Qualifiers**

X: Not ORELAP accredited.

Approved Signatory

Derrick Tanner  
General Manager


 Pixis Labs  
 Canrabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Abamectin	0.100	CIPC	1.000	Endrin	0.100
Acephate	0.100	Clethodim	0.050	EPN	0.050
Acequinocyl	0.100	Clethodim Sulfone	0.050	EPIC	0.100
Acetamiprid	0.020	Clethodim Sulfoxide	0.050	Esfenvalerate/ Fenvalerate	0.200
Acetochlor	0.100	Cbfezintezine	0.020	Etaconazole	0.100
Acrinathrin	0.100	Cbmazone	0.020	Ethalfuralin	0.100
Alachlor	0.100	Cbthianidin	0.200	Ethiofencarb	0.050
Aldicarb	0.100	Caumaphos	0.050	Ethion	0.200
Aldicarb sulfoxide	0.100	Crdoxyphos	0.020	Ethirimol	0.100
Aldoxycarb (Aldicarb-sulfone)	0.100	Cyazazine	0.020	Ethofumesate	0.050
Aldrin	0.100	Cyazofenphos	0.020	Ethoprophos	0.020
Ametoctradin	0.020	Cyatranylprole	0.050	Etofenprox	0.020
Ametryn	0.500	Cyazofamid	0.020	Etoxazole	0.020
Aspon	0.100	Cydoate	0.100	Etridiazole	0.100
Asulam	0.100	Cyfluthrin	0.200	Etrimefos	0.020
Atrazine	0.100	Cyhalothrin, lambda	0.200	Famoxadone	0.200
Atrazine-desethyl	0.100	Cymoxanil	0.050	Famphur	0.100
Azinphos-ethyl	0.020	Cypermethrin	0.200	Fenamidone	0.020
Azinphos-methyl	0.020	Cyprodinil	0.100	Fenamiphos	0.020
Azoxystrobin	0.020	Dadthal	0.100	Fenamiphos sulfone	0.020
Beralaxyl	0.020	Damhozide	0.100	Fenamiphos sulfoxide	0.020
Berdicarb	0.020	DCPMU	0.050	Fenazaquin	0.100
Berfluralin	0.100	DD, qp'-	0.100	Fenbuconazole	0.100
Berxacor	0.050	DD, p,p'-	0.100	Fenchlorphos	0.100
Bersulide	0.050	DDE, o,p'-	0.100	Fenchlorphos-oxon	0.100
BHC alpha isomer	0.100	DDE, p,p'-	0.100	Fenhexamid	0.100
BHC beta isomer	0.100	DDT, o,p'-	0.100	Fenitrothion	0.100
BHC delta isomer	0.500	DDT, p,p'-	0.100	Fenobucarb	0.050
Bifenazate	0.020	DEF (Tribufos)	0.100	Fenoxycarb	0.020
Bifenthrin	0.020	Deltamethrin	0.100	Fenpropathrin	0.050
Boscalid	0.020	Desmedipham	0.100	Fenpyroximate	0.020
Bromophos-ethyl	0.100	Diallate	0.100	Fenson	0.100
Bromophos-methyl	0.200	Diazinon	0.020	Fensulfthion	0.020
Bromopropylate	0.100	Diazoxon	0.100	Fensulfthion oxon	0.020
Bromuconazole	0.100	Dichlobenil	0.100	Fensulfthion sulfone	0.100
Bupirimate	0.020	Dichlofluanid	0.100	Fensulfthion-oxon-sulfone	0.020
Buprofezin	0.050	Dichlorvos	0.100	Fenthion	0.050
Butachlor	0.500	Diclobutrazol	0.050	Fenthion oxon	0.020
Butralin	0.200	Dicofol	0.100	Fenthion oxon sulfone	0.100
Butylate	0.100	Dicrotophos	0.050	Fenthion oxon sulfoxide	0.020
Cadusafos	0.020	Dieldrin	0.100	Fenthion sulfoxide	0.100
Captaf	1.000	Diethofencarb	0.020	Fenthion sulfone	0.050
Cartaryl	0.050	Diethyltoluamide (DEET)	0.050	Fenuron	0.020
Cartendazim	0.100	Difenoconazole	0.100	Fipronil	0.100
Cartofuran	0.020	Dimethenamid	0.050	Flonicamid	0.100
Cartophenothion	0.100	Dimethoate	0.050	Fluchloralin	0.100
Cartoxin	0.020	Dimethomorph	0.020	Flucythrinate	0.100
Carfentrazone-ethyl	0.100	Diniconazole	0.200	Fludioxonil	0.200
Chlorantraniliprole	0.020	Dinotefuran	0.200	Flufenacet	0.020
Chordane, cis-	0.200	Dioxathion	0.100	Rumioxazin	0.100
Chordane, trans-	0.200	Diphenamid	0.020	Ruometuron	0.020
Chlorfenapyr	0.500	Diphenylamine	0.100	Ruopicolide	0.050
Chlorfenson	0.200	Disulfoton	0.100	Fuopyram	0.020
Chlorfenvinphos	0.050	Disulfoton sulfone	0.100	Fuoxastrobin	0.050
Chlorobenzilate	0.100	Disulfoton sulfoxide	0.100	Fupyradifurone	0.020
Chloroneb	0.200	Diuron	0.050	Fluridone	0.100
Chlorpyrifos	0.050	Edifenphos	0.050	Fusilazole	0.020
Chlorpyrifos-methyl	0.200	Endosulfan apha	0.200	Futolanil	0.020
		Endosulfan beta	0.200	Futriafol	0.020
		Endosulfan sulfate	0.100	Fluvalinate, tau-	0.100
				Fluxapyroxad	0.020



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794

**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10



Fixis Labs  
Canabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)
Fomesafen	0.100	Mexacarbate	0.020	Propamocarb	0.050
Fonofos	0.100	MGK 264	0.020	Proparil	0.050
Forchlorfenuron	0.050	Mirex	0.100	Propargite	0.050
Formetanate	0.050	Molinate	0.050	Propazine	0.020
Furathiocarb	0.020	Monocrotophos	0.100	Propetamphos	0.050
Heptachlor	0.100	Monolinuron	0.020	Propham	0.050
Heptachlor epoxide	0.100	Myclobutanil	0.050	Propiconazole	0.050
Heptenophos	0.100	Naled	0.100	Propoxur	0.050
Hexachlorobenzene	0.100	Napropamide	0.050	Propoxycarbazone Na	0.050
Hexaconazole	0.100	Neburon	0.020	Propyzamide	0.050
Hexazinone	0.100	Nitrapyrin	0.100	Prthiofos	0.100
Hexythiazox	0.020	Norflurazon	0.050	Pyraclostrobin	0.020
Imazali	0.100	Omethoate	0.100	Pyrazophos	0.050
Imidacoprid	0.100	O-Phenylphenol	0.100	Pyrethrins	0.050
Indaziflam	0.020	Oxadixyl	0.100	Pyridaben	0.020
Indoxacarb	0.020	Oxaryl	0.100	Pyridafol	0.100
Iprobenfos	0.100	Oxamyl-oxime	0.100	Pyridate	0.020
Iprodione	0.100	Oxychlorane	0.100	Pyrimethanil	0.050
Isobenzan	0.100	Oxydemeton-Methyl	0.100	Pyriproxifen	0.020
Isocarbophos	0.500	Oxythioquinox	0.200	Pyroxasulfone	0.020
Isodrin	0.100	Padbutrazol	0.050	Pyrosulam	0.020
Isufenphos	0.050	Paraoxon-ethyl	0.020	Quinalphos	0.050
Isufenphos-methyl	0.020	Paraoxon methyl	0.100	Quinoxyfen	0.050
Isufenphos oxon	0.050	Parathion ethyl	0.100	Quintozene (PQNB)	0.200
Isoprocarb	0.020	Parathion methyl	0.200	Resmethrin	0.050
Isopropalin	0.200	Perconazole	0.050	Rotenone	0.050
Isoprothiolane	0.050	Perdimethalin	0.050	S421	0.100
Isoproturon	0.050	Perflufen	0.020	Smaizine	0.100
Isoxaben	0.050	Pertachloroaniline	0.100	Smectryl	0.200
Isoxaflutole	0.050	Pertachloroanisole	0.100	Spinetoram	0.020
Kresoxim-methyl	0.050	Pentachlorobenzene (PCB)	0.100	Spinosad	0.050
Lactofen	0.500	Pentachlorothiobenzene (PCTA)	0.100	Spirodiclofen	0.100
Lenacl	0.100	Perthiopyrad	0.020	Spiromesifen	0.050
Lindane (gammaBHC)	0.100	Permethrin	0.050	Spirotetramat	0.050
Linuron	0.020	Peithane	0.100	Spiroxamine	0.020
Malaaxon	0.050	Phenmedipharm	0.050	Sulfotep	0.050
Malathion	0.050	Phenthoate	0.050	Sulfoxaflor	0.050
Mandipropamid	0.020	Phorate	0.050	Sulprofos	0.020
Mecarbam	0.020	Phorate Sulfone	0.050	Tebuconazole	0.100
Mepanipyrim	0.050	Phorate Sulfoxide	0.050	Tebufenozide	0.020
Merphos	0.500	Phosalone	0.050	Tebuthiuron	0.020
Metalaxyl	0.050	Phosmet	0.100	Teconazene	0.100
Metaldehyde	0.050	Phosphamidon	0.050	Tefluthrin	0.100
Metconazole	0.100	Phoxim	0.050	Terbufos	0.020
Methacifos	0.100	Pinoxaden	0.020	Terbufos sulfone	0.050
Methamidophos	0.050	Piperonyl butoxide	0.050	Terbufos sulfoxide	0.050
Methidathion	0.050	Pirimicarb	0.020	Terbutylazine	0.020
Methiocarb	0.050	Pirimiphos-methyl	0.050	Terbutryn	0.020
Methiocarb sulfone	0.100	Pirimiphos-ethyl	0.020	Tetrachlorvinphos	0.050
Methiocarb sulfoxide	0.100	Prallethrin	0.100	Tetraconazole	0.050
Methomyl	0.100	Prochloraz	0.020	Tetradfon	0.200
Methoxychlor	0.100	Procyimdone	0.100	Tetramethrin	0.050
Methoxyfenozide	0.020	Prdenofos	0.100	Tetrasul	0.100
Metobromuron	0.050	Prfluralin	0.100	Thiabendazole	0.100
Metolachlor	0.100	Promecarb	0.050	Thiabendazole, 5-hydroxy	0.100
Metolcarb	0.050	Prometon	0.100	Thiadoprid	0.050
Metrifenone	0.050	Prometryn	0.020	Thiamethoxam	0.100
Metribuzin	0.100	Propadchlor	0.020	Thiobencarb	0.050
Mevinphos	0.100			Thiodicarb	0.050
				Thiophanate-methyl	0.050



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10



PIXIS Labs  
 Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)
Tolclofos-methyl	0.100	Triazophos	0.020	Trifloxystrobin	0.020
Triforin	0.100	Tolyfluarid	0.050	Triticonazole	0.050
Tralkoxydim	0.100	Tridiphane	0.500	Vindozolin	0.100
Triadimefon	0.050	Triflumizole	0.020	Zoxamide	0.020
Triallate	0.100	Trifluralin	0.100		

LOQ=Limit of Quantitation, mg/kg

Factors affecting the LOQ include instrumentation sensitivity for a particular analyte, sample size, moisture content (percent solids) of the sample, effectiveness of the cleanup on the sample extract, and especially the type of sample matrix.



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10



12423 NE Whitaker Way Portland OR 97230 p.503-254-1794

**Cannabis Chain of Custody Record**

ORELAP ID: OR100028

Company: NW Natural Goods		Analysis Requested											Purchase Order Number:				
Contact: Tyson Humble		Pesticides – OR 59 compounds	Pesticide Multi-Residue – 379 compounds	Potency	Residual Solvents	Water Activity	Moisture	Terpenes	Micro: Yeast and Mold	Micro: E. Coli and Total Coliform	Heavy Metals	Mycotoxins	Other	Full panel			Project Number:
Address: 11791 SE HWY 212, Clackamas, OR 97015														<input type="checkbox"/> Report Instructions: <input type="checkbox"/> Send to State - METRC <input checked="" type="checkbox"/> Email Final Results: <input type="checkbox"/> Fax Final Results <input type="checkbox"/> Cash/Check/CC/Net 30	Project Name:		
Email: tysonh@nwconfections.com															Other:		
Phone: 314-779-6557 Fax:																	
Processor's License: 330-1058115IHH														Field ID	Date/Time Collected	Matrix	Weight
1	BEV - BO 021021-1	1/26/21		X	X	X	X	X	X	X	X			water	362.1g	362.1g	

Collected By:	Relinquished By:	Date	Time	Received by:	Date	Time	Lab Use Only:
<input checked="" type="checkbox"/> Standard (5 day)	Isaac Velasquez	1/26/21	1045	[Signature]	1.27.21	1045	Client Alias: NWN Order Number: 21-000940
<input type="checkbox"/> Rush (3-4 day) (1.5x Standard)	[Signature]	1.27.21	1110	Magit Alan	01/27/21	10:30	Proper Container Sample Condition Temperature: 13.3 c
<input type="checkbox"/> Priority Rush (2 day) (2x Standard)							Shipped Via: Courier Evidence of cooling: <input type="checkbox"/> Yes <input type="checkbox"/> No

SUBMISSION OF SAMPLES WITH TESTING REQUIREMENTS TO PIXIS WILL BE UNDERSTOOD TO BE AN AGREEMENT FOR SERVICES IN ACCORDANCE WITH THE CONDITIONS LISTED ON THE BACK OF THIS FORM  
Revision: 1.00 Control#: CF023 Effective 11/8/2018 Revised 11/8/2018 www.pixislabs.com Page 1 of 2

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made. Testing in accordance with: OAR 333-007-0390 OAR 333-007-0400 OAR 333-007-0410 OAR 333-007-0430



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



Report Number: 21-000940/D02.R00  
 Report Date: 02/02/2021  
 ORELAP#: OR100028  
 Purchase Order:  
 Received: 01/27/21 11:10

Laboratory Quality Control Results									
Residual Solvents					Batch ID: 2100841				
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		660	595	µg/g	110.9	70	- 130
Isobutane	ND	< 200		804	761	µg/g	105.7	70	- 130
Butane	ND	< 200		845	761	µg/g	111.0	70	- 130
2,2-Dimethylpropane	ND	< 200		1020	955	µg/g	106.8	70	- 130
Methanol	ND	< 200		1580	1610	µg/g	98.1	70	- 130
Ethylene Oxide	ND	< 30		60.3	58.3	µg/g	103.4	70	- 130
2-Methylbutane	ND	< 200		1540	1630	µg/g	94.5	70	- 130
Pentane	ND	< 200		1430	1610	µg/g	88.8	70	- 130
Ethanol	ND	< 200		1460	1620	µg/g	90.1	70	- 130
Ethyl Ether	ND	< 200		1510	1630	µg/g	92.6	70	- 130
2,2-Dimethylbutane	ND	< 30		174	176	µg/g	98.9	70	- 130
Acetone	ND	< 200		1560	1610	µg/g	96.9	70	- 130
2-Propanol	ND	< 200		1510	1620	µg/g	93.2	70	- 130
Ethyl Formate	ND	< 500		1610	1610	µg/g	100.0	70	- 130
Acetonitrile	ND	< 100		418	499	µg/g	83.8	70	- 130
Methyl Acetate	ND	< 500		1650	1620	µg/g	101.9	70	- 130
2,3-Dimethylbutane	ND	< 30		135	165	µg/g	81.8	70	- 130
Dichloromethane	ND	< 200		411	482	µg/g	85.7	70	- 130
2-Methylpentane	ND	< 30		143	164	µg/g	87.2	70	- 130
MTBE	ND	< 500		1570	1600	µg/g	98.1	70	- 130
3-Methylpentane	ND	< 30		144	163	µg/g	88.3	70	- 130
Hexane	ND	< 30		140	166	µg/g	84.3	70	- 130
1-Propanol	ND	< 500		1440	1600	µg/g	90.0	70	- 130
Methylethylketone	ND	< 500		1460	1600	µg/g	91.3	70	- 130
Ethyl acetate	ND	< 200		1400	1610	µg/g	87.0	70	- 130
2-Butanol	ND	< 200		1370	1620	µg/g	84.6	70	- 130
Tetrahydrofuran	ND	< 100		441	509	µg/g	86.6	70	- 130
Cyclohexane	ND	< 200		1510	1620	µg/g	93.2	70	- 130
2-methyl-1-propanol	ND	< 500		1400	1600	µg/g	87.5	70	- 130
Benzene	ND	< 1		26.6	30.2	µg/g	88.1	70	- 130
Isopropyl Acetate	ND	< 200		1610	1610	µg/g	100.0	70	- 130
Heptane	ND	< 200		1600	1610	µg/g	99.4	70	- 130
1-Butanol	ND	< 500		1520	1610	µg/g	94.4	70	- 130
Propyl Acetate	ND	< 500		1700	1600	µg/g	106.3	70	- 130
1,4-Dioxane	ND	< 100		480	493	µg/g	97.4	70	- 130
2-Ethoxyethanol	ND	< 30		157	168	µg/g	93.5	70	- 130
Methylisobutylketone	ND	< 500		1740	1610	µg/g	108.1	70	- 130
3-Methyl-1-butanol	ND	< 500		1400	1610	µg/g	87.0	70	- 130
Ethylene Glycol	ND	< 200		339	483	µg/g	70.2	70	- 130
Toluene	ND	< 200		488	491	µg/g	99.4	70	- 130
Isobutyl Acetate	ND	< 500		1820	1610	µg/g	113.0	70	- 130
1-Pentanol	ND	< 500		1360	1610	µg/g	84.5	70	- 130
Butyl Acetate	ND	< 500		1790	1630	µg/g	109.8	70	- 130
Ethylbenzene	ND	< 200		974	966	µg/g	100.8	70	- 130
m,p-Xylene	ND	< 200		957	982	µg/g	97.5	70	- 130
o-Xylene	ND	< 200		988	976	µg/g	101.2	70	- 130
Cumene	ND	< 30		159	162	µg/g	98.1	70	- 130
Anisole	ND	< 500		1480	1620	µg/g	91.4	70	- 130
DMSO	ND	< 500		1310	1620	µg/g	80.9	70	- 130
1,2-dimethoxyethane	ND	< 50		150	172	µg/g	87.2	70	- 130
Triethylamine	ND	< 500		1330	1600	µg/g	83.1	70	- 130
N,N-dimethylformamide	ND	< 150		478	483	µg/g	99.0	70	- 130
N,N-dimethylacetamide	ND	< 150		493	499	µg/g	98.8	70	- 130
Pyridine	ND	< 50		173	166	µg/g	104.2	70	- 130



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

QC - Sample Duplicate Sample ID: 21-000905-0001

Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methylethylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	

**Abbreviations**  
 ND - None Detected at or above MRL  
 RPD - Relative Percent Difference  
 LOQ - Limit of Quantitation

**Units of Measure:**  
 µg/g - Microgram per gram or ppm



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

Revision #: 0.00 Control : CFL-D06  
Revision Date: 05/31/2019 Effective Date: 05/31/2019

Laboratory Quality Control Results

JAOAC2015 V98-6 Batch ID: 2100943

Laboratory Control Sample

Analyte	Result	Spike	Units	% Rec	Limits	Evaluation	Notes
CBDV-A	0.000913	0.001	%	91.3	85.0 - 115	Acceptable	
CBDV	0.000951	0.001	%	95.1	85.0 - 115	Acceptable	
CBD-A	0.000962	0.001	%	96.2	85.0 - 115	Acceptable	
CBG-A	0.000923	0.001	%	92.3	85.0 - 115	Acceptable	
CBG	0.000963	0.001	%	96.3	85.0 - 115	Acceptable	
CBD	0.000967	0.001	%	96.7	85.0 - 115	Acceptable	
THCV	0.000970	0.001	%	97.0	85.0 - 115	Acceptable	
THCVA	0.000868	0.001	%	86.8	85.0 - 115	Acceptable	
CBN	0.00107	0.001	%	107	85.0 - 115	Acceptable	
THC	0.000965	0.001	%	96.5	85.0 - 115	Acceptable	
D8THC	0.000983	0.001	%	98.3	85.0 - 115	Acceptable	
CBL	0.000904	0.001	%	90.4	85.0 - 115	Acceptable	
CBC	0.00101	0.001	%	101	85.0 - 115	Acceptable	
THCA	0.000966	0.001	%	96.6	85.0 - 115	Acceptable	
CBCA	0.000882	0.001	%	88.2	85.0 - 115	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDV-A	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBDV	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBD-A	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBG-A	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBG	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBD	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCV	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCVA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBN	<LOQ	0.0001	%	< 0.0001	Acceptable	
THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
D8THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBL	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBC	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBCA	<LOQ	0.0001	%	< 0.0001	Acceptable	

**Abbreviations**

ND - None Detected at or above MRL  
RPD - Relative Percent Difference  
LOQ - Limit of Quantitation

**Units of Measure:**

% - Percent



12423 NE Whitaker Way  
Portland, OR 97230  
503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

Revision #: 0.00 Control : CFL-D06  
Revision Date: 05/31/2019 Effective Date: 05/31/2019

Laboratory Quality Control Results

JAOAC2015 V98-6		Batch ID: 2100943						
Sample Duplicate		Sample ID: 21-000940-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDV-A	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBD-A	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBG-A	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBG	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBD	0.00718	0.00696	0.0001	%	3.13	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
THC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
D8THC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBCA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	

**Abbreviations**

- ND - None Detected at or above MRL
- RPD - Relative Percent Difference
- LOQ - Limit of Quantitation
- NA - Calculation Not Applicable given non-numerical results

**Units of Measure:**

% - Percent



12423 NE Whitaker Way  
 Portland, OR 97230  
 503-254-1794



**Report Number:** 21-000940/D02.R00  
**Report Date:** 02/02/2021  
**ORELAP#:** OR100028  
**Purchase Order:**  
**Received:** 01/27/21 11:10

Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.