

CONSOLIDATED TEST RESULTS SUMMARY

Please see the following pages for full test results.

BULK SKU	BATCH #	LOQ: Limit Of Quantitation	
PRODUCT NAME	SERVING SIZE	LOD: Limit Of Detection	
LABORATORY :	OREGON ACCREDITATION: OR100028	1 g = 10 ⁻³ kg = 10 ³ mg = 10 ⁶ µg 1 mg/kg = 1 ppm = 1000 ppb	
POTENCY	PER SERVING	PER GRAM	Percent
Cannabidiol (CBD)	mg/serving	mg/g	%
Total THC (d9-THC, THCA)	mg/serving	mg/g	%
Cannabigerol (CBG)	mg/serving	mg/g	%
Cannabinol (CBN)	mg/serving	mg/g	%
Cannabichromene (CBC)	mg/serving	mg/g	%
Tetrahydrocannabinolic Acid (THCA)	mg/serving	mg/g	%
Delta-9-THC (d9-THC)	mg/serving	mg/g	%
Delta-8-THC (d8-THC)	mg/serving	mg/g	%
HEAVY METALS	PER SERVING	PER GRAM	REGULATORY ACTION LEVEL
Arsenic	µg/serving	µg/g	10 µg/day ^[1]
Cadmium	µg/serving	µg/g	4.1 µg/day ^[1]
Lead	µg/serving	µg/g	6 µg/day ^[1]
Mercury	µg/serving	µg/g	2 µg/day ^[1]
PESTICIDES	REGULATORY ACTION LEVEL		
None of the other 59 pesticides tested found above limit of detection in the sample.			10 ppb ^[1]
RESIDUAL SOLVENTS	Results	REGULATORY ACTION LEVEL	
Ethanol*	µg/g	50,000 mg/day	
Heptane	µg/g	50,000 mg/day	
None of the 34 residual solvents tested found above limit of quantitation in the sample.			
MICROBIAL	PASS/FAIL		
Yeast & Mold	Pass		
Coliform	Pass		



1. American Herbal Pharmacopoeia. (2014). Cannabis Inflorescence: Standards of Identity, Analysis, and Quality Control. Washington DC: AHP.

*Ethanol is a food additive used in some of our ingredients. The FDA has labeled ethanol as Generally Recognized as Safe (GRAS). Many foods contain trace amounts of ethanol, including soy sauce, pasta sauces, fruits and juices, etc. Our products contain safe levels of ethanol and always below pertinent regulatory action levels.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Customer: Etz Hayim Holdings
Product identity: CYCL-SLZ.D9.GF5.6pk-EK29Z
Client/Metric ID: .
Laboratory ID: 23-000138-0008

Summary

Potency:

Analyte per 1g	Result	Limits	Units	Status	
CBD per 1g	0.0204		mg/1g		CBD-Total per Serving Size 0.0204 mg/1g
Δ8-THC per 1g	0.00219		mg/1g		
Δ9-THC per 1g	0.0111		mg/1g		THC-Total per Serving Size 0.0111 mg/1g
(Reported in milligrams per serving)					

Residual Solvents:

Analyte	Result (µg/g)	Limits (µg/g)	Status
Ethanol	1220		

Pesticides:

All analytes passing and less than LOQ.

Metals:

Less than LOQ for all analytes.

Microbiology:

Less than LOQ for all analytes.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Customer: Etz Hayim Holdings
 16427 NE Airport Way
 PORTLAND 97230
 United States of America (USA)
Product identity: CYCL-SLZ.D9.GF5.6pk-EK29Z
Client/Metric ID: .
Sample Date:
Laboratory ID: 23-000138-0008
Evidence of Cooling: No
Temp: 18.8 °C
Relinquished by: client
Serving Size #1: 1 g

Sample Results

Potency per 1g		Method: J AOAC 2015 V98-6 (mod) ^b		Units mg/se Batch: 2300237		Analyze: 1/7/23 4:00:00 AM	
Analyte	Result	Limits	Units	LOQ	Notes		
CBC per 1g	< LOQ		mg/1g	0.000962			
CBC-A per 1g	< LOQ		mg/1g	0.000962			
CBC-Total per 1g	< LOQ		mg/1g	0.00181			
CBD per 1g	0.0204		mg/1g	0.000962			
CBD-A per 1g	< LOQ		mg/1g	0.000962			
CBD-Total per 1g	0.0204		mg/1g	0.00181			
CBDV per 1g	< LOQ		mg/1g	0.000962			
CBDV-A per 1g	< LOQ		mg/1g	0.000962			
CBDV-Total per 1g	< LOQ		mg/1g	0.00180			
CBE per 1g	< LOQ		mg/1g	0.000962			
CBG per 1g	< LOQ		mg/1g	0.000962			
CBG-A per 1g	< LOQ		mg/1g	0.000962			
CBG-Total per 1g	< LOQ		mg/1g	0.00180			
CBL per 1g	< LOQ		mg/1g	0.000962			
CBL-A per 1g	< LOQ		mg/1g	0.000962			
CBL-Total per 1g	< LOQ		mg/1g	0.00181			
CBN per 1g	< LOQ		mg/1g	0.000962			
CBT per 1g	< LOQ		mg/1g	0.000962			
Δ8-THCV per 1g	< LOQ		mg/1g	0.000962			
Δ10-THC-9R per 1g	< LOQ		mg/1g	0.000962			
Δ8-THC per 1g	0.00219		mg/1g	0.000962			
Δ9-THC per 1g	0.0111		mg/1g	0.000962			
exo-THC per 1g	< LOQ		mg/1g	0.000962			
THC-A per 1g	< LOQ		mg/1g	0.000962			
THC-Total per 1g	0.0111		mg/1g	0.00181			
THCV per 1g	< LOQ		mg/1g	0.000962			
THCV-A per 1g	< LOQ		mg/1g	0.000962			
THCV-Total per 1g	< LOQ		mg/1g	0.00181			
Total Cannabinoids per 1g	0.0338		mg/1g				



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
E.coli	< LOQ		cfu/g	10	2300180	01/09/23 AOAC 991.14 (Petrifilm) ^P		
Total Coliforms	< LOQ		cfu/g	10	2300180	01/09/23 AOAC 991.14 (Petrifilm) ^P		
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2300181	01/10/23 AOAC 2014.05 (RAPID) ^P		
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2300181	01/10/23 AOAC 2014.05 (RAPID) ^P		

Solvents **Method:** Residual Solvents by GC/MS^b **Units** µg/g **Batch** 2300543 **Analyze** 01/18/23 10:44 AM

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethanol	1220		200		
Ethyl acetate	< LOQ	5000	200	pass		Ethyl benzene	< LOQ		200		
Ethyl ether	< LOQ	5000	200	pass		Ethylene glycol	< LOQ	620	200	pass	
Ethylene oxide	< LOQ	50.0	20.0	pass		Hexanes (sum)	< LOQ	290	150	pass	
Isopropyl acetate	< LOQ	5000	200	pass		Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass	
m,p-Xylene	< LOQ		200			Methanol	< LOQ	3000	200	pass	
Methylene chloride	< LOQ	600	60.0	pass		Methylpropane (Isobutane)	< LOQ		200		
n-Butane	< LOQ		200			n-Heptane	< LOQ	5000	200	pass	
n-Hexane	< LOQ		30.0			n-Pentane	< LOQ		200		
o-Xylene	< LOQ		200			Pentanes (sum)	< LOQ	5000	600	pass	
Propane	< LOQ	5000	200	pass		Tetrahydrofuran	< LOQ	720	100	pass	
Toluene	< LOQ	890	100	pass		Total Xylenes	< LOQ		400		
Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass							



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Pesticides											
Method: AOAC 2007.01 & EN 15662 (mod) ^b											
Units mg/kg Batch 2300273 Analyze 01/10/23 10:44 AM											
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin [‡]	< LOQ	0.50	0.250	pass		Acephate [‡]	< LOQ	0.40	0.200	pass	
Acequinocyl [‡]	< LOQ	2.0	1.00	pass		Acetamiprid [‡]	< LOQ	0.20	0.100	pass	
Aldicarb [‡]	< LOQ	0.40	0.200	pass		Azoxystrobin [‡]	< LOQ	0.20	0.100	pass	
Bifenazate [‡]	< LOQ	0.20	0.100	pass		Bifenthrin [‡]	< LOQ	0.20	0.100	pass	
Boscalid [‡]	< LOQ	0.40	0.200	pass		Carbaryl [‡]	< LOQ	0.20	0.100	pass	
Carbofuran [‡]	< LOQ	0.20	0.100	pass		Chlorantraniliprole [‡]	< LOQ	0.20	0.100	pass	
Chlorfenapyr [‡]	< LOQ	1.0	0.500	pass		Chlorpyrifos [‡]	< LOQ	0.20	0.100	pass	
Clofentezine [‡]	< LOQ	0.20	0.100	pass		Cyfluthrin [‡]	< LOQ	1.0	0.500	pass	
Cypermethrin [‡]	< LOQ	1.0	0.500	pass		Daminozide [‡]	< LOQ	1.0	0.500	pass	
Diazinon [‡]	< LOQ	0.20	0.100	pass		Dichlorvos [‡]	< LOQ	1.0	0.500	pass	
Dimethoate [‡]	< LOQ	0.20	0.100	pass		Ethoprophos [‡]	< LOQ	0.20	0.100	pass	
Etofenprox [‡]	< LOQ	0.40	0.200	pass		Etoxazole [‡]	< LOQ	0.20	0.100	pass	
Fenoxycarb [‡]	< LOQ	0.20	0.100	pass		Fenpyroximate [‡]	< LOQ	0.40	0.200	pass	
Fipronil [‡]	< LOQ	0.40	0.200	pass		Flonicamid [‡]	< LOQ	1.0	0.400	pass	
Fludioxonil [‡]	< LOQ	0.40	0.200	pass		Hexythiazox [‡]	< LOQ	1.0	0.400	pass	
Imazalil [‡]	< LOQ	0.20	0.100	pass		Imidacloprid [‡]	< LOQ	0.40	0.200	pass	
Kresoxim-methyl [‡]	< LOQ	0.40	0.200	pass		Malathion [‡]	< LOQ	0.20	0.100	pass	
Metalaxyl [‡]	< LOQ	0.20	0.100	pass		Methiocarb [‡]	< LOQ	0.20	0.100	pass	
Methomyl [‡]	< LOQ	0.40	0.200	pass		MGK-264 [‡]	< LOQ	0.20	0.100	pass	
Myclobutanil [‡]	< LOQ	0.20	0.100	pass		Naled [‡]	< LOQ	0.50	0.250	pass	
Oxamyl [‡]	< LOQ	1.0	0.500	pass		Pacllobutrazole [‡]	< LOQ	0.40	0.200	pass	
Parathion-Methyl [‡]	< LOQ	0.20	0.100	pass		Permethrin [‡]	< LOQ	0.20	0.100	pass	
Phosmet [‡]	< LOQ	0.20	0.100	pass		Piperonyl butoxide [‡]	< LOQ	2.0	1.00	pass	
Prallethrin [‡]	< LOQ	0.20	0.100	pass		Propiconazole [‡]	< LOQ	0.40	0.200	pass	
Propoxur [‡]	< LOQ	0.20	0.100	pass		Pyrethrin I (total) [‡]	< LOQ	1.0	0.500	pass	
Pyridaben [‡]	< LOQ	0.20	0.100	pass		Spinosad [‡]	< LOQ	0.20	0.100	pass	
Spiromesifen [‡]	< LOQ	0.20	0.100	pass		Spirotetramat [‡]	< LOQ	0.20	0.100	pass	
Spiroxamine [‡]	< LOQ	0.40	0.200	pass		Tebuconazole [‡]	< LOQ	0.40	0.200	pass	
Thiacloprid [‡]	< LOQ	0.20	0.100	pass		Thiamethoxam [‡]	< LOQ	0.20	0.100	pass	
Trifloxystrobin [‡]	< LOQ	0.20	0.100	pass							

Metals										
Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes		
Arsenic	< LOQ	0.200	mg/kg	0.00356	2300352	01/11/23 AOAC 2013.06 (mod.) ^b	pass			
Cadmium	< LOQ	0.200	mg/kg	0.00356	2300352	01/11/23 AOAC 2013.06 (mod.) ^b	pass			
Lead	< LOQ	0.500	mg/kg	0.00356	2300352	01/11/23 AOAC 2013.06 (mod.) ^b	pass			
Mercury	< LOQ	0.100	mg/kg	0.00178	2300352	01/11/23 AOAC 2013.06 (mod.) ^b	pass			



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Ⓟ = ISO/IEC 17025:2017 accredited method.

Ⓜ = TNI accredited analyte.

Units of Measure

cfu/g = Colony forming units per gram

g = g

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/1g = Milligram per 1g

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 Batch ID: 2300237

Laboratory Control Sample									
Analyte	LCS	Result	Spike	Units	% Rec	Limits		Evaluation	Notes
CBDVA	2	0.0010	0.001	%	95.9	80.0	- 120	Acceptable	
CBDV	2	0.0011	0.001	%	98.3	80.0	- 120	Acceptable	
CBE	2	0.0010	0.001	%	97.7	80.0	- 120	Acceptable	
CBDA	1	0.0010	0.001	%	96.5	90.0	- 110	Acceptable	
CBGA	1	0.0010	0.001	%	98.0	80.0	- 120	Acceptable	
CBG	1	0.0010	0.001	%	99.1	80.0	- 120	Acceptable	
CBD	1	0.0010	0.001	%	98.6	90.0	- 110	Acceptable	
THCV	2	0.0011	0.001	%	100	80.0	- 120	Acceptable	
d8THCV	2	0.0010	0.001	%	97.4	80.0	- 120	Acceptable	
THCVA	2	0.0009	0.001	%	94.3	80.0	- 120	Acceptable	
CBN	1	0.0011	0.001	%	99.6	80.0	- 120	Acceptable	
exo-THC	2	0.0010	0.001	%	98.7	80.0	- 120	Acceptable	
d9THC	1	0.0012	0.001	%	106	90.0	- 110	Acceptable	
d8THC	1	0.0010	0.001	%	95.8	90.0	- 110	Acceptable	
CBL	2	0.0010	0.001	%	103	80.0	- 120	Acceptable	
d10THC	1	NA	0.001	%	NA	80.0	- 120	Acceptable	Q6
CBG	2	0.0011	0.001	%	97.6	80.0	- 120	Acceptable	
THCA	1	0.0010	0.001	%	96.5	90.0	- 110	Acceptable	
CBCA	2	0.0010	0.001	%	92.8	80.0	- 120	Acceptable	
CBLA	2	0.0010	0.001	%	94.9	80.0	- 120	Acceptable	
CBT	2	0.0011	0.001	%	96.9	80.0	- 120	Acceptable	

Method Blank						
Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBDV	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBE	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBDA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBGA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBG	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBD	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCV	<LOQ	0.0001	%	< 0.0001	Acceptable	
d8THCV	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCVA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBN	<LOQ	0.0001	%	< 0.0001	Acceptable	
exo-THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
d9THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
d8THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBL	<LOQ	0.0001	%	< 0.0001	Acceptable	
d10THC	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBG	<LOQ	0.0001	%	< 0.0001	Acceptable	
THCA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBCA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBLA	<LOQ	0.0001	%	< 0.0001	Acceptable	
CBT	<LOQ	0.0001	%	< 0.0001	Acceptable	

Abbreviations
 ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:
 % - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 Batch ID: 2300237

Sample Duplicate Sample ID: 23-000124-0001

Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBD A	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBGA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBG	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBD	0.0158	0.0158	0.0001	%	0.502	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
d9THC	0.0088	0.0090	0.0001	%	2.08	< 20	Acceptable	
d8THC	0.0020	0.0019	0.0001	%	6.63	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
d10THC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBC	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBCA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.0001	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662		Units: mg/Kg			Batch ID: 2300273			
Method Blank		Laboratory Control Sample						
Analyte	Blank Result	Blank Limits	Notes	LCS Result	LCS Spike	LCS % Rec	Limits	Notes
Abamectin	0.000	< 0.250		1.003	1.000	100.3	50.0	150
Acephate	0.000	< 0.200		0.767	0.800	95.8	60.0	120
Acetaminophen	0.000	< 1.000		3.914	4.000	97.9	40.0	160
Acetamiprid	0.000	< 0.100		0.388	0.400	97.1	60.0	120
Aldicarb	0.000	< 0.200		0.834	0.800	104.3	60.0	120
Azoxystrobin	0.000	< 0.100		0.381	0.400	95.3	60.0	120
Bifenazate	0.000	< 0.100		0.429	0.400	107.4	60.0	120
Bifenthrin	0.000	< 0.100		0.397	0.400	99.2	50.0	150
Boscalid	0.000	< 0.200		0.780	0.800	97.5	60.0	120
Carbaryl	0.000	< 0.100		0.394	0.400	98.5	60.0	120
Carbofuran	0.000	< 0.100		0.391	0.400	97.7	60.0	120
Chlorantraniliprole	0.000	< 0.100		0.397	0.400	99.2	60.0	120
Chlorfenapyr	0.000	< 0.500		1.974	2.000	98.7	60.0	120
Chlorpyrifos	0.000	< 0.100		0.370	0.400	92.6	60.0	120
Clofentazine	0.000	< 0.100		0.289	0.400	72.1	60.0	120
Cyfluthrin	0.000	< 0.500		1.965	2.000	98.2	50.0	150
Cypermethrin	0.000	< 0.500		1.915	2.000	95.8	50.0	150
Daminozide	0.000	< 0.500		0.674	2.000	33.7	60.0	120
Diazinon	0.000	< 0.100		0.410	0.400	102.5	60.0	120
Dichlorvos	0.000	< 0.500		1.953	2.000	97.6	60.0	120
Dimethoate	0.000	< 0.100		0.403	0.400	100.7	60.0	120
Ethoprophos	0.000	< 0.100		0.392	0.400	98.0	60.0	120
Etofenprox	0.000	< 0.200		0.760	0.800	95.0	50.0	150
Etoxazole	0.000	< 0.100		0.400	0.400	100.1	60.0	120
Fenoxycarb	0.000	< 0.100		0.389	0.400	97.1	60.0	120
Fenpyroximate	0.000	< 0.200		0.767	0.800	95.9	60.0	120
Fipronil	0.000	< 0.200		0.814	0.800	101.8	60.0	120
Fonicamid	0.000	< 0.250		1.026	1.000	102.6	60.0	120
Fludioxonil	0.000	< 0.200		0.807	0.800	100.9	50.0	150
Hexythiazox	0.000	< 0.250		0.961	1.000	96.1	60.0	120
Imazalil	0.000	< 0.100		0.402	0.400	100.6	60.0	120
Imidacloprid	0.000	< 0.200		0.796	0.800	99.4	60.0	120
Kresoxim-methyl	0.000	< 0.200		0.803	0.800	100.4	60.0	120
Malathion	0.000	< 0.100		0.392	0.400	98.1	60.0	120
Metaxalyl	0.000	< 0.100		0.400	0.400	100.0	60.0	120
Methiocarb	0.000	< 0.100		0.399	0.400	99.8	60.0	120
Methomyl	0.000	< 0.200		0.819	0.800	102.4	60.0	120
MGK-264	0.000	< 0.100		0.395	0.400	98.6	50.0	150
Myclobutanil	0.000	< 0.100		0.398	0.400	99.6	60.0	120
Naled	0.000	< 0.250		0.984	1.000	98.4	50.0	150
Oxamyl	0.000	< 0.500		2.004	2.000	100.2	60.0	120
Paclotrazole	0.000	< 0.200		0.783	0.800	97.9	60.0	120
Parathion-Methyl	0.000	< 0.100		0.409	0.400	102.2	50.0	150
Permethrin	0.000	< 0.100		0.383	0.400	95.8	50.0	150
Phosmet	0.000	< 0.100		0.401	0.400	100.3	50.0	150
Piperonyl butoxide	0.000	< 0.500		2.064	2.000	103.2	60.0	120
Prallethrin	0.000	< 0.100		0.402	0.400	100.4	60.0	120
Propiconazole	0.000	< 0.200		0.815	0.800	101.9	60.0	120
Propoxur	0.000	< 0.100		0.399	0.400	99.9	60.0	120
Pyrethrin (Summe)	0.001	< 0.100		0.490	0.488	100.4	60.0	120
Pyridaben	0.000	< 0.100		0.386	0.400	96.6	50.0	150
Spinosad	0.000	< 0.100		0.381	0.388	98.2	50.0	150
Spiromesifen	0.000	< 0.100		0.383	0.400	95.9	60.0	120
Spirotetramat	0.000	< 0.100		0.385	0.400	96.2	60.0	120
Spiroxamine	0.000	< 0.200		0.788	0.800	98.5	60.0	120
Tebuconazole	0.000	< 0.200		0.800	0.800	100.0	60.0	120
Thiacloprid	0.000	< 0.100		0.395	0.400	98.8	60.0	120
Thiamethoxam	0.000	< 0.100		0.414	0.400	103.6	60.0	120
Trifloxystrobin	0.000	< 0.100		0.393	0.400	98.2	60.0	120

Q6



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662		Units: mg/Kg					Batch ID: 2300273			
Matrix Spike/Matrix Spike Duplicate Recoveries		Sample ID: 23-000025-0001								
Analyte	Result	MS Res	MSD Res	Spike	RPD%	Limit	MS % Rec	MSD % Rec	Limits	Notes
Abamectin	0.000	0.938	0.979	1.000	4.2%	< 30	93.8%	97.9%	50 - 150	
Acephate	0.000	0.746	0.743	0.800	0.4%	< 30	93.3%	92.9%	50 - 150	
Acetaminopyr	0.000	3.577	3.754	4.000	4.8%	< 30	89.4%	93.9%	50 - 150	
Acetamiprid	0.000	0.383	0.383	0.400	0.2%	< 30	95.9%	95.7%	50 - 150	
Aldicarb	0.000	0.809	0.823	0.800	1.7%	< 30	101.2%	102.8%	50 - 150	
Azoxystrobin	0.000	0.369	0.371	0.400	0.5%	< 30	92.2%	92.6%	50 - 150	
Bifenazate	0.000	0.386	0.397	0.400	2.8%	< 30	96.5%	99.3%	50 - 150	
Bifenthrin	0.000	0.362	0.373	0.400	3.1%	< 30	90.5%	93.3%	50 - 150	
Boscalid	0.000	0.758	0.766	0.800	1.1%	< 30	94.8%	95.8%	50 - 150	
Carbaryl	0.000	0.377	0.391	0.400	3.6%	< 30	94.3%	97.7%	50 - 150	
Carbofuran	0.000	0.368	0.380	0.400	3.1%	< 30	92.0%	94.9%	50 - 150	
Chlorantraniliprole	0.000	0.375	0.391	0.400	4.1%	< 30	93.8%	97.8%	50 - 150	
Chlorfenapyr	0.000	1.900	1.861	2.000	2.1%	< 30	95.0%	93.1%	50 - 150	
Chlorpyrifos	0.000	0.333	0.345	0.400	3.4%	< 30	83.4%	86.2%	50 - 150	
Clofentezine	0.000	0.054	0.052	0.400	3.8%	< 30	13.4%	12.9%	50 - 150	Q
Cyfluthrin	0.000	1.975	2.108	2.000	6.5%	< 30	98.8%	105.4%	30 - 150	
Cypermethrin	0.000	1.931	2.012	2.000	4.1%	< 30	96.5%	100.6%	50 - 150	
Daminozide	0.000	0.650	0.672	2.000	3.3%	< 30	32.5%	33.6%	30 - 150	
Diazinon	0.000	0.324	0.341	0.400	5.1%	< 30	81.0%	85.2%	50 - 150	
Dichlorvos	0.000	1.890	1.963	2.000	3.8%	< 30	94.5%	98.1%	50 - 150	
Dimethoate	0.000	0.381	0.389	0.400	1.9%	< 30	95.3%	97.2%	50 - 150	
Ethoprophos	0.000	0.375	0.376	0.400	0.1%	< 30	93.8%	93.9%	50 - 150	
Etofenprox	0.000	0.709	0.728	0.800	2.6%	< 30	88.6%	91.0%	50 - 150	
Etoxazole	0.000	0.363	0.375	0.400	3.4%	< 30	90.6%	93.8%	50 - 150	
Fenoxycarb	0.000	0.368	0.373	0.400	1.3%	< 30	92.0%	93.2%	50 - 150	
Fenpyroximate	0.000	0.808	0.826	0.800	2.3%	< 30	101.0%	103.3%	50 - 150	
Fipronil	0.000	0.865	0.903	0.800	4.3%	< 30	108.1%	112.9%	50 - 150	
Fonicamid	0.000	1.014	1.030	1.000	1.6%	< 30	101.4%	103.0%	50 - 150	
Fludioxonil	0.000	0.765	0.758	0.800	0.9%	< 30	95.6%	94.8%	50 - 150	
Hexythiazox	0.000	0.831	0.824	1.000	0.9%	< 30	83.1%	82.4%	50 - 150	
Imazalil	0.000	0.379	0.390	0.400	2.9%	< 30	94.8%	97.6%	50 - 150	
Imidacloprid	0.000	0.739	0.764	0.800	3.4%	< 30	92.4%	95.5%	50 - 150	
Kresoxim-methyl	0.000	0.740	0.761	0.800	2.8%	< 30	92.5%	95.2%	50 - 150	
Malathion	0.000	0.370	0.383	0.400	3.4%	< 30	92.6%	95.8%	50 - 150	
Metaxalaxyl	0.000	0.380	0.382	0.400	0.4%	< 30	95.1%	95.5%	50 - 150	
Methiocarb	0.000	0.379	0.386	0.400	1.9%	< 30	94.7%	96.5%	50 - 150	
Methomyl	0.000	0.817	0.811	0.800	0.8%	< 30	102.1%	101.3%	50 - 150	
MGK-264	0.000	0.369	0.369	0.400	0.1%	< 30	92.2%	92.3%	50 - 150	
Myclobutanil	0.000	0.366	0.388	0.400	5.8%	< 30	91.5%	97.0%	50 - 150	
Naled	0.000	0.930	0.963	1.000	3.5%	< 30	93.0%	96.3%	50 - 150	
Oxamyl	0.000	1.968	2.071	2.000	5.1%	< 30	98.4%	103.5%	50 - 150	
Paclobutrazole	0.000	0.755	0.776	0.800	2.7%	< 30	94.4%	97.0%	50 - 150	
Parathion-Methyl	0.000	0.384	0.407	0.400	6.0%	< 30	95.9%	101.8%	30 - 150	
Permethrin	0.000	0.363	0.385	0.400	5.9%	< 30	90.6%	96.2%	50 - 150	
Phosmet	0.000	0.378	0.375	0.400	0.6%	< 30	94.4%	93.8%	50 - 150	
Piperonyl butoxide	0.000	1.815	1.959	2.000	7.6%	< 30	90.8%	98.0%	50 - 150	
Prallethrin	0.000	0.369	0.382	0.400	3.6%	< 30	92.2%	95.6%	50 - 150	
Propiconazole	0.000	0.751	0.766	0.800	1.9%	< 30	93.9%	95.8%	50 - 150	
Propoxur	0.000	0.379	0.389	0.400	2.6%	< 30	94.7%	97.2%	50 - 150	
Pyrethrin (Summe)	0.000	0.631	0.643	0.488	1.9%	< 30	129.3%	131.7%	50 - 150	
Pyridaben	0.000	0.369	0.372	0.400	1.0%	< 30	92.2%	93.1%	50 - 150	
Spirosad	0.000	0.365	0.379	0.388	3.7%	< 30	94.1%	97.6%	50 - 150	
Spiromesifen	0.000	0.360	0.372	0.400	3.3%	< 30	89.9%	93.0%	50 - 150	
Spirotetramat	0.000	0.367	0.371	0.400	1.0%	< 30	91.8%	92.8%	50 - 150	
Spiroxamine	0.000	0.749	0.773	0.800	3.2%	< 30	93.6%	96.7%	50 - 150	
Tebuconazole	0.000	0.762	0.767	0.800	0.8%	< 30	95.2%	95.9%	50 - 150	
Thiacloprid	0.000	0.380	0.390	0.400	2.4%	< 30	95.1%	97.4%	50 - 150	
Thiamethoxam	0.000	0.410	0.410	0.400	0.1%	< 30	102.5%	102.5%	50 - 150	
Trifloxystrobin	0.000	0.336	0.351	0.400	4.3%	< 30	84.1%	87.8%	50 - 150	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-000138/D002.R000
Report Date: 01/19/2023
ORELAP#: OR100028
Purchase Order:
Received: 01/04/23 15:10





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.